

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/098
 Appliance Serial Number: xxxx
 Report Creation Time: P661DT13H17M46S
 Upper alarm limit Above +8.0 C for 10h
 Lower alarm limit Below -0.5 C for 1h

Day	Average storage temp.(C)	Status	Low Temperature			High Temperature			Door open count	Door open time (hh:mm)	AC Power avail. (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp.(C)	Faults
			Min. temp. (C)	Total time below 2 C (hh:mm)	Total low alarm time (hh:mm)	Max. Temp. (C)	Total time above 8 C (hh:mm)	Total high alarm time (hh:mm)						
today	2.1	OK	1.8	9:12	0:00	2.7	0:00	0:00	3	0:00	2:23	1:38	24.2	D
2	2.0	OK	1.8	5:03	0:00	2.1	0:00	0:00	0	0:00	5:43	4:17	24.0	D
3	2.0	OK	1.7	12:00	0:00	2.1	0:00	0:00	0	0:00	5:22	2:08	25.5	D
4	1.9	OK	1.6	22:12	0:00	2.0	0:00	0:00	2	0:00	5:46	1:44	25.2	D
5	1.9	OK	1.6	16:30	0:00	2.1	0:00	0:00	0	0:00	5:10	2:18	25.4	D
6	1.9	OK	1.7	14:33	0:00	2.1	0:00	0:00	0	0:00	5:42	4:20	24.6	D
7	1.9	OK	1.7	17:57	0:00	2.1	0:00	0:00	0	0:00	5:08	1:47	25.6	D
8	1.8	OK	1.5	20:02	0:00	2.1	0:00	0:00	0	0:00	5:31	3:03	25.0	D
9	2.1	OK	1.9	6:27	0:00	2.1	0:00	0:00	0	0:00	5:30	2:23	25.0	D
10	1.8	OK	1.5	20:45	0:00	2.0	0:00	0:00	0	0:00	5:16	2:38	24.8	D
11	1.9	OK	1.6	17:03	0:00	2.3	0:00	0:00	0	0:00	5:23	3:32	25.0	D
12	2.2	OK	1.6	11:57	0:00	2.8	0:00	0:00	2	0:00	4:22	1:40	25.1	D
13	1.9	OK	1.7	13:50	0:00	2.1	0:00	0:00	0	0:00	5:36	4:42	24.5	D
14	1.9	OK	1.6	21:24	0:00	2.0	0:00	0:00	0	0:00	5:54	1:48	25.1	D
15	1.9	OK	1.6	19:00	0:00	2.0	0:00	0:00	0	0:00	5:19	2:28	25.4	D
16	1.8	OK	1.7	17:51	0:00	2.1	0:00	0:00	0	0:00	5:47	4:59	24.5	D
17	2.1	OK	2.0	0:15	0:00	2.2	0:00	0:00	0	0:00	5:22	2:32	24.5	D
18	1.9	OK	1.7	15:54	0:00	2.1	0:00	0:00	0	0:00	5:09	1:38	25.5	D
19	1.8	OK	1.6	14:06	0:00	2.0	0:00	0:00	0	0:00	3:05	0:38	24.8	D A
20	1.8	OK	1.7	5:36	0:00	2.0	0:00	0:00	0	0:00	0:00	0:00	27.0	D A
21	2.0	OK	2.0	1:59	0:00	2.1	0:00	0:00	0	0:00	1:16	0:00	27.3	D
22	1.9	OK	1.7	3:18	0:00	2.2	0:00	0:00	0	0:00	0:19	1:23	26.4	D A
23	1.8	OK	1.6	8:02	0:00	2.1	0:00	0:00	0	0:00	2:23	1:50	27.1	D A
24	1.8	OK	1.7	5:28	0:00	2.3	0:00	0:00	0	0:00	1:16	0:19	26.7	D A
25	1.7	OK	1.6	2:40	0:00	2.0	0:00	0:00	0	0:00	0:26	0:17	26.5	D A
26	2.2	OK	1.9	0:12	0:00	2.3	0:00	0:00	0	0:00	1:07	1:07	24.9	D A
27	1.8	OK	1.7	2:48	0:00	1.8	0:00	0:00	0	0:00	0:00	0:00	27.4	D A
28	1.7	OK	1.7	3:19	0:00	1.8	0:00	0:00	0	0:00	0:00	0:00	26.9	D A
29	2.0	OK	1.9	2:35	0:00	2.1	0:00	0:00	0	0:00	1:32	0:34	27.3	D A
30	1.9	OK	1.6	4:17	0:00	2.1	0:00	0:00	0	0:00	0:58	2:17	26.9	D A
31	2.0	OK	1.7	1:58	0:00	2.1	0:00	0:00	0	0:00	0:58	1:02	26.8	D A
32	1.7	OK	1.7	6:23	0:00	1.9	0:00	0:00	0	0:00	0:57	0:54	27.5	D A

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/098
 Appliance Serial Number: xxxx
 Report Creation Time: P661DT13H17M46S
 Upper alarm limit: Above +8.0 C for 10h
 Lower alarm limit: Below -0.5 C for 1h

Day	Average storage temp.(C)	Status	Low Temperature			High Temperature			Door open count	Door open time (hh:mm)	AC Power avail. (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp.(C)	Faults
			Min. temp. (C)	Total time below 2 C (hh:mm)	Total low alarm time (hh:mm)	Max. Temp. (C)	Total time above 8 C (hh:mm)	Total high alarm time (hh:mm)						
33														
34	1.8	OK	1.7	6:42	0:00	1.8	0:00	0:00	0	0:00	1:18	0:20	27.0	D A
35	1.8	OK	1.7	5:18	0:00	1.9	0:00	0:00	0	0:00	0:42	0:00	27.2	D A
36	1.7	OK	1.6	5:35	0:00	2.0	0:00	0:00	0	0:00	0:49	0:48	27.2	D A
37	2.0	OK	2.0	0:30	0:00	2.1	0:00	0:00	0	0:00	2:01	0:00	27.0	D A
38	1.8	OK	1.6	7:01	0:00	2.1	0:00	0:00	0	0:00	2:08	2:10	27.0	D A
39	2.0	OK	1.9	0:09	0:00	2.1	0:00	0:00	0	0:00	1:06	0:08	27.1	D A
40	1.9	OK	1.8	2:15	0:00	2.3	0:00	0:00	2	0:00	0:01	0:03	27.5	D A
41	1.9	OK	1.7	5:04	0:00	2.1	0:00	0:00	0	0:00	2:23	1:53	27.2	D A
42	2.0	OK	2.0	0:15	0:00	2.1	0:00	0:00	0	0:00	0:25	0:25	25.1	D A
43	1.9	OK	1.8	5:09	0:00	2.1	0:00	0:00	0	0:00	0:58	1:08	27.6	D A
44	1.8	OK	1.7	2:26	0:00	1.8	0:00	0:00	0	0:00	0:00	0:00	27.8	D A
45	1.7	OK	1.6	4:38	0:00	1.8	0:00	0:00	0	0:00	1:17	1:12	27.4	D A
46	2.0	OK	1.9	0:20	0:00	2.1	0:00	0:00	0	0:00	0:36	0:26	25.2	D A
47	1.9	OK	1.8	5:30	0:00	2.1	0:00	0:00	0	0:00	2:11	1:21	27.0	D A
48	2.0	OK	2.0	1:06	0:00	2.0	0:00	0:00	0	0:00	1:04	0:39	24.5	D A
49	1.8	OK	1.8	2:52	0:00	2.0	0:00	0:00	0	0:00	0:23	0:23	27.0	D A
50	2.1	OK	1.8	0:02	0:00	2.1	0:00	0:00	0	0:00	0:32	0:32	24.8	D A
51	2.1	OK	1.9	0:01	0:00	2.1	0:00	0:00	0	0:00	0:33	0:33	24.6	D A
52														
53	2.1	OK	1.6	0:17	0:00	2.3	0:00	0:00	0	0:00	0:29	0:01	25.5	D A
54	2.1	OK	2.1	0:00	0:00	2.1	0:00	0:00	0	0:00	0:14	0:14	23.1	D A
55	1.9	OK	1.7	0:16	0:00	2.1	0:00	0:00	0	0:00	0:02	0:03	25.8	D A
56	1.7	OK	1.6	5:42	0:00	1.9	0:00	0:00	0	0:00	1:40	0:50	27.1	D A
57														
58	2.2	OK	1.9	0:09	0:00	2.5	0:00	0:00	2	0:00	0:36	0:36	25.1	D A
59														
60	1.9	OK	1.9	0:21	0:00	1.9	0:00	0:00	0	0:00	0:20	0:00	24.2	D A

Fault Key (System):

D - Freezer Temp. sensor unreadable

Fault Key (Compressor):

A - low battery