

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/098
 Appliance Serial Number: xxxx
 Report Creation Time: P587DT0H39M33S
 Upper alarm limit: Above +8.0 C for 10h
 Lower alarm limit: Below -0.5 C for 1h

Day	Average storage temp.(C)	Status	Low Temperature			High Temperature			Door open count	Door open time (hh:mm)	AC Power avail. (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp.(C)	Average Faults
			Min. temp. (C)	Total time below 2 C (hh:mm)	Total low alarm time (hh:mm)	Max. Temp. (C)	Total time above 8 C (hh:mm)	Total high alarm time (hh:mm)						
today	7.7	OK	7.7	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	0:00	22.7	
2	7.6	OK	7.4	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	0:00	21.9	
3	7.0	OK	5.9	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	0:34	22.2	
4	7.5	OK	6.7	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	6:34	23.2	
5	7.7	OK	7.5	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	3:15	22.7	
6	7.6	OK	7.5	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	1:57	22.3	
7	7.7	OK	7.5	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	2:38	22.9	
8	7.6	OK	7.4	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	0:00	23.6	
9	7.1	OK	5.6	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	0:00	23.4	
10	7.2	OK	5.7	0:00	0:00	7.6	0:00	0:00	0	0:00	0:00	3:36	22.5	
11	7.6	OK	7.4	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	3:25	22.3	
12	7.7	OK	7.5	0:00	0:00	8.3	2:30	0:00	0	0:00	0:00	4:55	22.4	
13	7.6	OK	7.5	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	0:00	23.0	
14	7.2	OK	5.7	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	0:00	23.7	
15	7.4	OK	5.9	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	4:16	22.8	
16	7.6	OK	7.3	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	4:15	22.2	
17	7.8	OK	7.4	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	3:27	22.3	
18	7.7	OK	7.6	0:00	0:00	8.0	0:15	0:00	0	0:00	0:00	0:00	23.4	
19	7.0	OK	5.6	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	0:30	23.4	
20	7.3	OK	6.5	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	3:49	22.6	
21	7.8	OK	7.2	0:00	0:00	8.0	0:30	0:00	0	0:00	0:00	2:57	23.7	A
22	7.8	OK	7.7	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	1:41	22.4	
23	7.5	OK	6.8	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	2:15	22.3	
24	7.6	OK	7.2	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	7:02	23.1	
25	7.7	OK	7.6	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	6:27	22.9	
26	7.9	OK	7.6	0:00	0:00	8.2	2:32	0:00	0	0:00	0:00	5:44	22.9	
27	7.9	OK	7.7	0:00	0:00	8.2	7:45	0:00	0	0:00	0:00	5:05	22.9	
28	7.9	OK	7.8	0:00	0:00	8.1	6:00	0:00	0	0:00	0:00	0:57	22.3	
29	8.0	ALARM	7.8	0:00	0:00	8.3	9:15	10:30	0	0:00	0:00	0:00	22.2	
30	8.0	ALARM	7.9	0:00	0:00	8.1	7:43	0:17	0	0:00	0:00	0:00	22.1	
31	8.0	ALARM	7.9	0:00	0:00	8.2	17:00	14:58	0	0:00	0:00	0:00	22.4	
32	7.9	ALARM	7.7	0:00	0:00	8.1	5:17	1:15	0	0:00	0:00	0:15	22.3	

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/098
 Appliance Serial Number: xxxx
 Report Creation Time: P587DT0H39M33S
 Upper alarm limit: Above +8.0 C for 10h
 Lower alarm limit: Below -0.5 C for 1h

Day	Average storage temp.(C)	Status	Low Temperature			High Temperature			Door open count	Door open time (hh:mm)	AC Power avail. (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp.(C)	Faults
			Min. temp. (C)	Total time below 2 C (hh:mm)	Total low alarm time (hh:mm)	Max. Temp. (C)	Total time above 8 C (hh:mm)	Total high alarm time (hh:mm)						
33	8.0	ALARM	7.8	0:00	0:00	8.2	12:45	3:30	0	0:00	0:00	1:57	22.7	
34	7.9	OK	7.7	0:00	0:00	8.1	1:43	0:00	0	0:00	0:00	0:46	22.4	
35	7.9	OK	7.7	0:00	0:00	8.1	3:32	0:00	0	0:00	0:00	3:43	22.6	
36	7.9	ALARM	7.8	0:00	0:00	8.2	7:30	0:30	0	0:00	0:00	2:03	22.5	
37	7.9	OK	7.7	0:00	0:00	8.1	1:28	0:00	0	0:00	0:00	2:02	22.5	
38	7.8	OK	7.7	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	0:07	22.2	A
39	7.7	OK	7.6	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	0:28	23.0	A
40	7.3	OK	6.5	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	0:00	23.2	A
41	6.7	OK	5.5	0:00	0:00	7.5	0:00	0:00	0	0:00	0:00	3:59	22.9	A
42	7.5	OK	7.1	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	8:38	22.3	A
43	7.6	OK	7.4	0:00	0:00	7.6	0:00	0:00	0	0:00	0:00	3:49	21.9	A
44	7.7	OK	7.6	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	0:02	21.1	A
45	7.6	OK	7.4	0:00	0:00	7.6	0:00	0:00	0	0:00	0:00	0:23	21.2	A
46	7.7	OK	7.5	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	3:55	21.8	A
47	7.7	OK	7.6	0:00	0:00	8.0	0:15	0:00	0	0:00	0:00	2:33	21.8	A
48	7.6	OK	7.6	0:00	0:00	8.0	0:17	0:00	0	0:00	0:00	0:57	21.4	A
49	7.7	OK	7.6	0:00	0:00	8.0	0:43	0:00	0	0:00	0:00	3:33	21.9	A
50	7.6	OK	7.5	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	0:54	21.0	A
51	7.7	OK	7.5	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	1:11	21.3	A
52	7.8	OK	7.7	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	1:40	22.1	A
53	7.7	OK	7.5	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	0:00	22.7	A
54	6.9	OK	5.5	0:00	0:00	7.6	0:00	0:00	0	0:00	0:00	0:00	22.6	
55	7.2	OK	5.7	0:00	0:00	7.7	0:00	0:00	0	0:00	0:00	5:40	22.3	
56	7.7	OK	7.5	0:00	0:00	8.0	0:00	0:00	0	0:00	0:00	4:46	22.3	
57	7.6	OK	7.4	0:00	0:00	7.9	0:00	0:00	0	0:00	0:00	0:00	22.6	
58	6.9	OK	5.5	0:00	0:00	7.6	0:00	0:00	0	0:00	0:00	0:00	22.7	
59	6.9	OK	5.6	0:00	0:00	7.5	0:00	0:00	0	0:00	0:00	4:32	22.3	
60	7.5	OK	7.0	0:00	0:00	7.8	0:00	0:00	0	0:00	0:00	4:16	22.0	

Fault Key (System):

A - low battery

Fault Key (Compressor):