

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/xxx
 Appliance Serial Number: 102410
 Report Creation Time: P98DT21H40M22S
 Upper alarm limit: Above +8.0 C for 10h
 Lower alarm limit: Below -0.5 C for 1h

Day	Average storage temp.(C)	Status	Low Temperature			High Temperature			Door open count	Door open time (hh:mm)	AC Power avail. (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp.(C)
			Min. temp. (C)	Total time below 2 C (hh:mm)	Total low alarm time (hh:mm)	Max. Temp. (C)	Total time above 8 C (hh:mm)	Total high alarm time (hh:mm)					
today	8.1	ALARM	7.9	0:00	0:00	14.4	13:33	7:30	0	0:00	21:22	21:10	27.7
2	7.9	OK	7.6	0:00	0:00	10.4	5:58	0:00	1	0:00	23:52	23:53	27.9
3	7.3	OK	7.0	0:00	0:00	8.0	0:00	0:00	0	0:00	24:07	24:07	28.1
4	7.2	OK	7.0	0:00	0:00	7.3	0:00	0:00	0	0:00	24:00	24:00	28.5
5	7.1	OK	7.0	0:00	0:00	7.2	0:00	0:00	0	0:00	24:00	24:00	28.7
6	7.2	OK	7.0	0:00	0:00	7.3	0:00	0:00	0	0:00	24:00	24:00	29.0
7	7.0	OK	6.9	0:00	0:00	10.7	0:22	0:00	0	0:00	23:59	23:59	28.0
8	7.1	OK	6.8	0:00	0:00	13.4	0:38	0:00	2	0:00	23:38	23:37	26.8
9	7.0	OK	6.9	0:00	0:00	7.2	0:00	0:00	0	0:00	24:03	24:03	27.0
10	7.0	OK	6.8	0:00	0:00	7.2	0:00	0:00	0	0:00	23:59	23:59	26.8
11	7.0	OK	6.9	0:00	0:00	7.2	0:00	0:00	1	0:00	24:01	24:01	26.7
12	7.0	OK	6.9	0:00	0:00	7.1	0:00	0:00	0	0:00	24:01	24:01	26.7
13	7.0	OK	6.9	0:00	0:00	7.4	0:00	0:00	0	0:00	24:00	24:00	26.8
14	10.6	OK	6.8	0:00	0:00	22.9	6:10	0:00	1	0:00	19:05	19:05	26.1
15	25.2	ALARM	21.0	0:00	0:00	26.3	18:05	16:00	2	0:00	0:00	0:00	25.2
16	25.5	ALARM	24.7	0:00	0:00	26.3	24:00	24:00	0	0:00	0:00	0:00	25.5
17	24.9	ALARM	24.4	0:00	0:00	25.5	24:00	24:00	0	0:00	0:00	0:00	24.9
18	25.0	ALARM	24.5	0:00	0:00	25.6	24:00	24:00	0	0:00	0:00	0:00	25.0
19	25.0	ALARM	24.1	0:00	0:00	25.5	24:00	24:00	0	0:00	0:00	0:00	24.9
20	24.5	ALARM	24.2	0:00	0:00	25.1	24:00	24:00	0	0:00	0:00	0:00	24.5
21	24.6	ALARM	23.9	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	24.5
22	24.9	ALARM	24.5	0:00	0:00	25.4	24:00	24:00	0	0:00	0:00	0:00	24.9
23	24.9	ALARM	24.5	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	24.9
24	24.7	ALARM	23.9	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	24.6
25	25.7	ALARM	24.9	0:00	0:00	26.4	24:00	24:00	0	0:00	0:00	0:00	25.6
26	26.1	ALARM	25.7	0:00	0:00	26.6	24:00	24:00	0	0:00	0:00	0:00	26.1
27	25.3	ALARM	24.9	0:00	0:00	25.9	24:00	24:00	0	0:00	0:00	0:00	25.2
28	25.5	ALARM	24.6	0:00	0:00	26.3	24:00	24:00	0	0:00	0:00	0:00	25.4
29	25.0	ALARM	24.7	0:00	0:00	25.7	24:00	24:00	0	0:00	0:00	0:00	24.9
30	24.8	ALARM	24.3	0:00	0:00	25.1	24:00	24:00	0	0:00	0:00	0:00	24.8
31	25.4	ALARM	24.7	0:00	0:00	26.0	24:00	24:00	0	0:00	0:00	0:00	25.4
32	26.0	ALARM	25.5	0:00	0:00	26.3	24:00	24:00	0	0:00	0:00	0:00	26.0

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/xxx
 Appliance Serial Number: 102410
 Report Creation Time: P98DT21H40M22S
 Upper alarm limit: Above +8.0 C for 10h
 Lower alarm limit: Below -0.5 C for 1h

Day	Average storage temp.(C)	Status	Low Temperature			High Temperature			Door open count	Door open time (hh:mm)	AC Power avail. (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp.(C)
			Min. temp. (C)	Total time below 2 C (hh:mm)	Total low alarm time (hh:mm)	Max. Temp. (C)	Total time above 8 C (hh:mm)	Total high alarm time (hh:mm)					
33	25.8	ALARM	25.4	0:00	0:00	26.0	24:00	24:00	0	0:00	0:00	0:00	25.7
34	25.1	ALARM	24.7	0:00	0:00	25.5	24:00	24:00	0	0:00	0:00	0:00	25.0
35	25.1	ALARM	24.5	0:00	0:00	25.8	24:00	24:00	0	0:00	0:00	0:00	25.1
36	25.0	ALARM	24.6	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	25.0
37	25.4	ALARM	24.9	0:00	0:00	25.8	24:00	24:00	0	0:00	0:00	0:00	25.4
38	25.5	ALARM	24.8	0:00	0:00	26.2	24:00	24:00	0	0:00	0:00	0:00	25.5
39	24.7	ALARM	24.3	0:00	0:00	26.2	24:00	24:00	0	0:00	0:00	0:00	24.7
40	24.8	ALARM	24.1	0:00	0:00	25.4	24:00	24:00	0	0:00	0:00	0:00	24.8
41	25.8	ALARM	24.9	0:00	0:00	26.4	24:00	24:00	0	0:00	0:00	0:00	25.7
42	24.7	ALARM	23.6	0:00	0:00	25.9	24:00	24:00	0	0:00	0:00	0:00	24.6
43	24.2	ALARM	23.9	0:00	0:00	24.8	24:00	24:00	0	0:00	0:00	0:00	24.2
44	24.2	ALARM	23.3	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	24.2
45	24.5	ALARM	23.4	0:00	0:00	25.4	24:00	24:00	0	0:00	0:00	0:00	24.4
46	25.3	ALARM	24.5	0:00	0:00	26.0	24:00	24:00	0	0:00	0:00	0:00	25.2
47	24.9	ALARM	24.5	0:00	0:00	25.7	24:00	24:00	0	0:00	0:00	0:00	24.8
48	24.5	ALARM	24.0	0:00	0:00	25.0	24:00	24:00	0	0:00	0:00	0:00	24.5
49	24.5	ALARM	23.6	0:00	0:00	25.0	24:00	24:00	0	0:00	0:00	0:00	24.4
50	24.7	ALARM	24.4	0:00	0:00	25.0	24:00	24:00	0	0:00	0:00	0:00	24.7
51	25.1	ALARM	24.5	0:00	0:00	25.6	24:00	24:00	0	0:00	0:00	0:00	25.0
52	25.0	ALARM	24.2	0:00	0:00	25.5	24:00	24:00	0	0:00	0:00	0:00	24.9
53	24.3	ALARM	23.4	0:00	0:00	25.4	24:00	24:00	0	0:00	0:00	0:00	24.3
54	25.0	ALARM	23.3	0:00	0:00	26.7	24:00	24:00	0	0:00	0:00	0:00	24.8
55	24.2	ALARM	23.7	0:00	0:00	25.6	24:00	24:00	0	0:00	0:00	0:00	24.1
56	24.5	ALARM	24.1	0:00	0:00	24.9	24:00	24:00	0	0:00	0:00	0:00	24.4
57	24.5	ALARM	23.7	0:00	0:00	24.9	24:00	24:00	0	0:00	0:00	0:00	24.4
58	24.3	ALARM	23.4	0:00	0:00	24.9	24:00	24:00	0	0:00	0:00	0:00	24.2
59	24.3	ALARM	23.4	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	24.2
60	24.5	ALARM	23.7	0:00	0:00	25.3	24:00	24:00	0	0:00	0:00	0:00	24.4

Fault Key (System):

Fault Key (Compressor):